


# **TEMPORARY RULES COMMITTEE RULES REVIEW BOOK**

**Submitted for Review Before  
Senate Judiciary & Rules Committee  
65th Idaho Legislature  
First Regular Session – 2019**



*Prepared by:*

*Office of the Administrative Rules Coordinator  
Department of Administration*

*January 2019*

***SENATE JUDICIARY & RULES COMMITTEE***

***ADMINISTRATIVE RULES REVIEW***

***Table of Contents***

***2019 Legislative Session***

***IDAPA 11 – IDAHO STATE POLICE***  
***Idaho State Forensic Laboratory***

***11.03.01 – Rules Governing Alcohol Testing***

Docket No. 11-0301-1801 .....3

**IDAPA 11 – IDAHO STATE POLICE  
IDAHO STATE FORENSIC LABORATORY**

**11.03.01 – RULES GOVERNING ALCOHOL TESTING**

**DOCKET NO. 11-0301-1801**

**NOTICE OF RULEMAKING – ADOPTION OF TEMPORARY RULE**

**THE FOLLOWING NOTICE PUBLISHED WITH THE TEMPORARY RULE**

**EFFECTIVE DATE:** The effective date of the temporary rule is December 14, 2017.

**AUTHORITY:** In compliance with Sections 67-5226, Idaho Code, notice is hereby given this agency has adopted a temporary rule. The action is authorized pursuant to Section 67-2901(6), Idaho Code.

**DESCRIPTIVE SUMMARY:** The following is the required finding and concise statement of its supporting reasons for adopting a temporary rule:

This rule is being amended to update the incorporation by reference. The National Highway Traffic Safety Administration (NHTSA) updated the conforming products list (CPL) on November 2, 2017. Idaho State Police has no notification when these updates will occur in the federal register. This new CPL posting in the federal register updates the former incorporation by reference in the IDAPA 11.03.01 document and clearly delineates that all the evidentiary breath testing instruments in Idaho are approved for use on the NHTSA CPL. A question has been raised by the courts in Idaho if the Lifeloc FC20 and FC20BT instruments are materially the same instrument. There is no difference in the instruments other than the Bluetooth ability. Idaho performed a validation of the FC20BT model to ensure that nothing impacted the quality of testing done in Idaho. The Lifeloc FC20 is listed by name on the 2010 CPL and the FC20BT does not formally appear by name on the list until the 2017 federal register posting. While ISPFS has all the documentation from NHTSA, DOT, Volpe, and Lifeloc Corporation (and our own validation study) to show that there was no change to the instrument to give it Bluetooth capability, and that the instrument was considered still on the CPL by NHTSA, the new incorporation by reference document makes it abundantly clear to the Idaho courts in a federal register document.

**TEMPORARY RULE JUSTIFICATION:** Pursuant to Sections 67-5226(1)(a) and (b), Idaho Code, the Governor has found that temporary adoption of the rule is appropriate for the following reasons:

Necessary to protect the public health, safety or welfare and to conform to changes in the federal regulatory code.

**FEE SUMMARY:** Pursuant to Section 67-5226(2), the Governor has found that the fee or charge being imposed or increased is justified and necessary to avoid immediate danger and the fee is described herein: N/A

**ASSISTANCE ON TECHNICAL QUESTIONS:** For assistance on technical questions concerning the temporary rule, contact Matthew Gamette, (208) 884-7217.

DATED this December 15, 2017.

Colonel Kedrick R. Wills  
Director  
Idaho State Police  
700 S. Stratford Dr.  
Meridian, ID 83642  
Phone: (208) 884-7003  
Fax: (208) 884-7290

**FOLLOWING IS THE TEXT OF THE TEMPORARY RULE FOR DOCKET NO. 11-0301-1801**

**004. INCORPORATION BY REFERENCE.**

The following are incorporated by reference in this chapter of rules:

(4-7-11)

**01. Conforming Products List of Evidential Breath Measurement Devices (revised ~~3/11/2010~~ **11/2/2017**).** This document is available on the Internet at <http://edocket.access.gpo.gov/2010/pdf/2010-5242.pdf> <https://www.gpo.gov/fdsys/pkg/FR-2017-11-02/pdf/2017-23869.pdf>.  
(~~4-7-11~~)(**12-14-17**)T